

Supplier Returns Detailed Report

CID 4614 - Maxtec Inc

Supplier Returns Not Sent

Box: 1068 | Status: Requested RMA | Created: 09/Feb/2026 | RMA:

Stock Ref	Supplier Ref	Description	Serial Number	Fault Code	Qty	Line Type
0110432	R125P02-003	Maxtec Oxygen sensor MAX-250(B)			2	TOTAL
0110432	R125P02-003		LC77899195	No Output Country of origin: USA Custom Value: 45.00 \$		DETAIL
0110432	R125P02-003		LE90799101	No Output Country of origin: USA Custom Value: 45.00 \$		DETAIL

Box: 1057 | Status: Requested RMA | Created: 14/Nov/2025 | RMA: 358012

Stock Ref	Supplier Ref	Description	Serial Number	Fault Code	Qty	Line Type
0110429	R125P03-002	Maxtec Oxygen sensor MAX-250E			1	TOTAL
0110429	R125P03-002		KH39399174	Non Linear Country of origin: USA Customs Value \$46.50		DETAIL
0110468	R140P07-001	Maxtec Oxygen Sensor MAX-125M			1	TOTAL
0110468	R140P07-001		LG36999137	Non Linear Country of origin: USA Customs Value \$68.72		DETAIL

Shipments Sent but not Completed