

DVM Calibration record

DVM Calibration Index Number	CE098
DVM Bar code	
Equipment Used	Time 1030 MicroCal
Serial number Test Equipment	6839k6
Calibration Certificate Number	1098426
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
Micro Cal @	DVM Reading <i>mv</i>
0mV +/- 0.1	0.00
5mV +/- 0.1	5
10mV +/- 0.1	10
15mV +/- 0.1	15
20mV +/- 0.1	20
25mV +/- 0.1	25
30mV +/- 0.1	30
35mV +/- 0.1	35
40mv +/- 0.1	40
Pass / Fail	PAS
Tested By	DLAMB
Tested On	28/9/12
Signed	<i>D Lamb</i>