

1030 MicroCal Calibration record Date 28/09/2012

Microcal Calibration Index Number	CE076
Microcal Bar code	102060
Equipment Used	Pico
Serial number Test Equipment	ZNR2850
Calibration Certificate Number	8898
Test Equipment Calibration Index Number	CE174 (Barcode : 406730)
Micro Cal @	DVM Reading

Micro Cal 0-100MV	Pico	Micro Cal 0-10MV	
0mV +/- 0.1	0.003	0mV +/- 0.01	0.003
5mV +/- 0.1	4.972	1mV +/- 0.01	0.997
10mV +/- 0.1	9.959	2mV +/- 0.01	2.001
15mV +/- 0.1	15.004	3mV +/- 0.01	3.006
20mV +/- 0.1	19.997	4mV +/- 0.01	4.005
25mV +/- 0.1	25.017	5mV +/- 0.01	5.003
30mV +/- 0.1	30.008	6mV +/- 0.01	6.005
35mV +/- 0.1	35.019	7mV +/- 0.01	7.006
		8mV +/- 0.01	8.006
		9mV +/- 0.01	9.000
		10mV +/- 0.01	9.995
Pass / Fail	PAS		
Tested By	DLAMB		
Tested On	28/9/2012		
Signed			