

DVM Calibration record

DVM Calibration Index Number	CE 112
DVM Bar code	
Equipment Used	Time 1030 MicroCal
Serial number Test Equipment	6839k6
Calibration Certificate Number	1098426
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
Micro Cal @	DVM Reading
0mV +/- 0.1	0.0
5mV +/- 0.1	5mV
10mV +/- 0.1	10mV
15mV +/- 0.1	15mV
20mV +/- 0.1	20mV
25mV +/- 0.1	25mV
30mV +/- 0.1	30mV
35mV +/- 0.1	35mV
40mV +/- 0.1	39.9mV
Pass / Fail	PASS
Tested By	DLAMB
Tested On	J/10/12
Signed	DLamb