

DVM Calibration record

DVM Calibration Index Number	CE055
DVM Bar code	
Equipment Used	Time 1030 MicroCal
Serial number Test Equipment	6839k6
Calibration Certificate Number	1098426
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
Micro Cal @	DVM Reading
0mV +/- 0.1	0.000
5mV +/- 0.1	0.049V
10mV +/- 0.1	0.099V
15mV +/- 0.1	0.149V
20mV +/- 0.1	0.199V
25mV +/- 0.1	0.249V
30mV +/- 0.1	0.299V
35mV +/- 0.1	0.349V
40mv +/- 0.1	0.399
Pass / Fail	PASS
Tested By	DLAMB
Tested On	11/10/12
Signed	D Lamb