

Service Repair Sheet 63520

Contact Name Philip Strickland  
Company/ Hospital Name Medivent  
Department  
Position  
Direct Phone  
General Phone 003 5316280338  
Opera Account 00007375  
Email philip@medivent.ie  
Order Number  
Date Received 12/Sep/2012  
Booked in By Jonathan Connor

---

Repair	Ref	S/N	Equipment Type	Unable To Repair	Warranty
SRN21516	0014704	EBA00065	Pulse Oximetry Finger Sensor	Time :0 Hour(s)	

Parts Replaced  
Qty: 1 0014704 SpO2 Silicone Finger Sensor - Adult.

Notes:

<p>Replaced with probe serial number ECB02457</p>

Engineer Notes:

<br>12/Sep/2012 Philip Crossley<br><p>sensor has broken wire at connector</p><br>12/Sep/2012 Philip Crossley<br><p>Repl

---

Repair	Ref	S/N	Equipment Type	Unable To Repair	Warranty
SRN21517	0014704	EAK01885	Pulse Oximetry Finger Sensor	Time :0 Hour(s)	

Parts Replaced  
Qty: 1 0014704 SpO2 Silicone Finger Sensor - Adult.

Notes:

<p>Replaced with probe serial number ECB02458</p>

Engineer Notes:

<br>12/Sep/2012 Philip Crossley<br><p>no led, possibly failed or broken wire</p><br>12/Sep/2012 Philip Crossley<br><p>Repl

---

Repair	Ref	S/N	Equipment Type	Unable To Repair	Warranty
SRN21518	0014704	EBA00049	Pulse Oximetry Finger Sensor	Time :0 Hour(s)	

Parts Replaced  
Qty: 1 0014704 SpO2 Silicone Finger Sensor - Adult.

Notes:

<p>Replaced with probe serial number ECB02459</p>

Engineer Notes:

<br>12/Sep/2012 Philip Crossley<br><p>sensor wire broken at connector</p><br>12/Sep/2012 Philip Crossley<br><p>Repl

---