

breathalyser

AlcoTrue® M Breathalyser

Professional breathalyser for rapid accurate determination of breath alcohol concentration for medical use

Ideal for use by

- Emergency Services including paramedics
- Rehabilitation Clinics
- Accident and Emergency Departments



Key Features

- Professional fuel cell technology
- Precision intelligent microprocessor sampling system
- Automatic self-test after switched on
- Rapid, precise and reliable measurements, even at sub-zero ambient conditions
- Intuitive user interface
- Ready to test < 6 seconds
- Results are shown on a colour OLED display
- Compact, ergonomic design
- Low maintenance and spare part costs
- PC software included

Professional Breathalyser for rapid accurate determination of breath alcohol concentration in exhaled air. AlcoTrue M is for diagnostic use, e.g. hospitals and rehabilitation clinics

AlcoTrue M Standard Package

- AlcoTrue M Breathalyser; with active & fast modes**
- 25 mouthpieces (individually wrapped)
- Battery holder (4 x AA)
- 4 x AA batteries
- Instructions for use
- Small hard carrying case

** Further modes can be activated by Viamed, please confirm when ordering should you require alternative modes



Key Features of the AlcoTrue M Breathalyser

- Professional fuel cell technology
- Precision intelligent microprocessor sampling system
- Automatic self-test after switch on
- Rapid, precise and reliable measurements, even at sub-zero ambient conditions
- Intuitive user interface
- Ready to test < 6 seconds
- Results are shown on a colour OLED display
- Compact, ergonomic design
- Low maintenance and spare part costs
- PC software included




Measurement Modes

The AlcoTrue M provided five measurement modes, ideal for any situation:

- Active mode AM
- Fast mode FM
- Passive mode PM
- Scan mode SM*
- Mouth alcohol mode MA*

* Measurement modes only available if activated.



Applications

The AlcoTrue Professional Breathalyser is ideal for use by:

- Accident and Emergency Departments
- Emergency Services including paramedics
- Rehabilitation Clinics

AlcoTrue M Breathalyser Kit

- AlcoTrue M Breathalyser
- Battery holder (4 x AA)
- 4 x AA batteries
- Li-Poly battery
- UK or EU* main power adapter with USB data cable
- Vehicle power adapter
- 25 mouthpieces (individually wrapped)
- AlcoPrint printer
- Printer cable
- AlcoPrint UK or EU* mains power cable
- AlcoPrint vehicle power cable
- Single roll of printer paper
- PC software & USB cable
- Large hard carry case

* Must be specified at time of ordering.



The AlcoPrint Printer

The **AlcoTrue M** Breathalyser Kit includes a AlcoPrint USB wireless printer, this can print wirelessly from a range of 5 - 10 metres or connected directly to the unit with the USB cable provided. The AlcoPrint printer is powered by a vehicle or mains* power cable (provided).

The Printer Set is also available to purchase separately and includes: AlcoPrint Printer, Printer Cable, UK or EU* Mains and Vehicle adapter and a roll of printer paper.



AlcoPrint USB Wireless Printer

PC Software Kit

The **AlcoTrue M** PC Software Kit includes a USB cable and PC software CD.

The PC Software CD is also available without the USB cable.



PC Software



Mouthpieces and AlcoPrint paper rolls

Additional Accessories

Mouthpieces are available in packs of 25.

AlcoPrint paper rolls are available in packs of 5, each roll allows approximately ??? printed readings.

Should you require single pieces please do not hesitate in contacting us.

AlcoTrue® M Breathalyser

Sensor type	A318 Electrochemical sensor				
Sampling systems	High precision sampling system 0.2ml				
Measurement modes	Active mode (default) Fast mode (activated) Passive mode (activated) Mouth alcohol mode (if activated) Scanning mode (if activated)				
Measurement range					
Breath alcohol concentration (BrAC)	0.00 to 2.6 mg/L, 0 to 2600 µL, 0.000 to 0.550 g/210L, 0 to 260 µg/100mL, 0 to 260 µg% (Ethanol mass per breath volume at 34°C and 1013 hPa.)				
Blood alcohol concentration (BAC)	0.00 to 5.50%, 0.000 to 0.550 %, 0.00 to 5.50 g/L 0 to 550 mg/100mL, 0 to 550 mg%, 0.00 to 5.50 mg/mL (Ethanol mass per blood volume resp. mass at 20°C abd 1013 hPa)				
This data is given using a conversion factor of 2100:1 between breath - and blood alcohol concentrations. Other converion factors are available upon request.					
Ambient conditions					
For operations	-5 to 50°C (23 to 122°F), 10 to 100 R.H., 600 to 1400 hPa				
For storage	-20 to 65°C (-4 to 149°F)				
Sample conditions					
Minimal acceptable flow	Adjustable upon request, 12 L/min (set as default)				
Minimal volume	Adjustable, 1.2L (set as default)				
Measurement accuracy					
Reproducibility with ethanol standard					
0 to 0.476 mg/L	0.008 mg/L	0 to 1%.	0.017 mg/L		
> 0.476 mg/l to 1 mg/L	4% of measured value	> 1 % to 2 %.	4% of measured value		
> 1 mg/L	8% of measured value	> 2 %.	8% of measured value		
Sensitivity drift of measured value/month	Typically 0.5%				
Ready for operation					
After switching on	<6 seconds				
After completing a measurement in range (breath alcohol unit):					
0.00 mg/L	≤ 2 seconds				
0.25 mg/L	≤ 4 seconds				
0.50 mg/L	≤ 6 seconds				
1.00 mg/L	≤ 10 seconds				
>1.5 mg/L	≤ 15 seconds				
After completing a measurement in range (blood alcohol unit):					
0.00%.	≤ 2 seconds				
0.50%.	≤ 4 seconds				
1.00%.	≤ 6 seconds				
2.00%.	≤ 10 seconds				
>3.00%.	≤ 15 seconds				
Dimensions (W x H x D)	185 mm x 55 mm x 40 mm				
Weight	240g (including Alkaline batteries) 195g (including Li-Poly battery)				
Display	OLED, 256,000 colours				
Data memory	Up to 9999 measurements				
Calibration required	12 months				

€ 0482

Distributed by:



Viamed Limited · 15 Station Road · Cross Hills
 Keighley · West Yorkshire · BD20 7DT · United Kingdom
 Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582
 Email: info@viamed.co.uk · Website: www.viamed.co.uk



Part Number: 4790020
 Date: 12/12



BS EN ISO 9001:2008
 ISO 13485:2003

Screen Displays

Qualitative results indications
 (Scanning and Mouth Alcohol modes)



Alcohol detected in gas sample



No alcohol detected in gas sample

Quantitative results indications
 (Active, Fast, Passive and
 Mouth alcohol modes)



Pass



Alert



Fail