

Date: Tue, 30 Sep 2014 13:36:42 +0100
From: Steve Hardaker <steve.hardaker@viamed.co.uk>
Subject: R-42V and Stephan 130060001 compatibility
To: Onye Obiorah <onye.obiorah@gstt.nhs.uk>
Mime-Version: 1.0
Organization: Viamed
X-Mailer: GoldMine [6.70.50123]

Dear Mr Obiorah,

I just wanted to update you with regards to an issue that we discussed some time ago concerning the Viamed R-42V sensor. You may recall that your experiences with this sensor led you to the conclusion that the sensor is too slow for use in the Stephan Stephanie vent.

The R-42V sensor has a T90 response time of <12s; you stated that the original Stephan 130060001 sensor appears to have a faster response time, however, we cannot find anything to substantiate that. Viamed are technically able to produce a faster sensor but this sensor was designed to be a compatible replacement for the Stephan 130060001 sensor, which we believe to have the same specified response time (although in practice, it appears that it may well exceed the specification).

We would like to be able to offer you a sensor that meets your requirements but due to the lack of a written specification stating a faster response time to use as a new benchmark, we currently are not in a position to change the specification of the R-42V.

I have looked in depth for such a specification but without success. I appreciate that your comments regarding the response of the Stephan sensor are genuine and based on your observations, but can I ask if you are able to point me in the right direction of anything in writing that might confirm that the 130060001 sensor has a faster specified response?

At Viamed we continually strive to improve our products and I am hopeful that we can find confirmation to allow us to get around this current obstacle. If you have any information to add, or any comments regarding this, please let me know. Thank you in advance.

Regards,

Steve Hardaker
UK Sales Manager
Viamed Ltd.
<http://www.viamed.co.uk>
email: steve.hardaker@viamed.co.uk
Tel: +44 (0)1535 634542
Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd
Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

The information in this email is confidential and may be legally privileged.
It is intended solely for the addressee. Access to this email by anyone else is unauthorized.
If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

[Sent via Goldmine]