



Weldwide Solutions Limited

WGA-01 Oxygen Sensor

Website: www.weldwidesolutions.com
Email: sales@weldwidesolutions.com
Phone: +44 (0)1554 700 350



2013-02



S/N WS1xxxxx



Weldwide Solutions Limited

Units 8 & 18
Delta Lakes Enterprise Centre
The Avenue, Delta Lakes
Llanelli, Carmarthenshire, SA15 2DR
Website: www.weldwidesolutions.com
Email: sales@weldwidesolutions.com
Phone: +44 (0)1554 700 350

WGA-01 Oxygen Sensor



2013-02

S/N WS1xxxxx



This oxygen sensor contains caustic liquid, in case of leakage avoid contact with eyes and skin. Do not puncture or staple this bag. Do not remove the oxygen sensor from this bag until ready to use. Depending upon the ambient conditions the oxygen sensor may require 30 minutes to reach signal stability after installation.