
From: Steve Nixon [mailto:steve.nixon@viamed.co.uk]
Sent: Tuesday, May 14, 2013 9:51 AM
To: Vasu P Narasimhan
Cc: Kunal Kothari; Vasu P Narasimhan
Subject: FWD: R-24MED quality issues

Dear Vasu

Do you have any update on this matter?

Regards

Steve

----- Forwarded Message -----

From: Steve Nixon <steve.nixon@viamed.co.uk>
To: Vasu Narasimhan <VNarasimhan@teledyne.com>
Date: Mon, 22 Apr 2013 11:00:20 +0100
Subject: R-24MED quality issues

Dear Vasu

There appears to be some quality issues regarding R-24MED sensors:

- a) Molex connector off centre and Molex pins bent to one side, (2 pictures attached)
- b) Input thread damaged, (1 picture attached)
- c) Scuffing of plastic casing.

Steve