



Derek Lamb <liquidgands@gmail.com>

FW: MHRA Ref: 2013/005/020/401/001 - SpiroTrue A flow sensors

1 message

Emma.Rooke@mhra.gsi.gov.uk <Emma.Rooke@mhra.gsi.gov.uk>

To: liquidgands@gmail.com

28 January 2014 15:44

Dear Derek,

Thank you for the update regarding the SpiroTrue A.

The investigation has now been reviewed in light of our correspondence and it has been decided not to pursue the matter further at this time. The information obtained so far will be retained and will be available should it be required to support any future investigation in this area. I will continue to monitor the situation and would welcome details of any additional or similar incidents.

Thank you for your assistance in this matter and for the information you have provided. Please contact me if you wish to discuss this or other matters.

Kind regards,
Emma

Emma Rooke
Medical Device Specialist - MHRA
Tel: 0203 080 6609

Date: 08/01/2014 15:32:00

From: liquidgands@gmail.com On Behalf Of
Derek Lamb

To Emma Rooke,

cc

bcc

Subject: FW: MHRA Ref: 2013/005/020/401/001 -
SpiroTrue A flow sensors

From: liquidgands@gmail.com [mailto:liquidgands@gmail.com] **On Behalf Of** Derek Lamb
Sent: 08 January 2014 15:32
To: Rooke, Emma
Subject: MHRA Ref: 2013/005/020/401/001 - SpiroTrue A flow sensors

Dear Emma,

For your information, We are now superseding the SpiroTrue A flow sensor, with the SpiroTrue APC flow sensor, (The autoclave version). The price barrier has now been removed and both are produced at the roughly same cost there is no reason to justify having 2 ranges.

This in essence removes the problem of if the device should be single patient use or not.

Regards

Derek Lamb
Viamed Ltd.

This email was received from the INTERNET and scanned by the Government Secure Intranet anti-virus service supplied by Vodafone in partnership with Symantec. (CCTM Certificate Number 2009/09/0052.) In case of problems, please call your organisation's IT Helpdesk.

Communications via the GSi may be automatically logged, monitored and/or recorded for legal purposes.

This email and any files transmitted with it are **confidential**. If you are not the intended recipient, any reading, printing, storage, disclosure, copying or any other action taken in respect of this email is prohibited and may be unlawful.

If you are not the intended recipient, please notify the sender immediately by using the reply function and then permanently delete what you have received. Incoming and outgoing email messages are routinely monitored for compliance with the Department of Healths policy on the use of electronic communications.

For more information on the Department of Healths email policy, click

http://www.dh.gov.uk/DHTermsAndConditions/fs/en?CONTENT_ID=4110945&chk=x1C3Zw

The original of this email was scanned for viruses by the Government Secure Intranet virus scanning service supplied by Vodafone in partnership with Symantec. (CCTM Certificate Number 2009/09/0052.) On leaving the GSi this email was certified virus free.

Communications via the GSi may be automatically logged, monitored and/or recorded for legal purposes.