

Subject : Flowsensor A
Date : Mon, 13 May 2013 15:21:00 +0100
Linked to : Frank Williams
From : Catrin Hollings <catrin.hollings@viamed.co.uk>
To : Frank Williams <frank.williams@dchft.nhs.uk>

Dear Frank,

Further to our telephone conversation please find attached Operator Instructions for the Flowsensor A. My apologies but I was incorrect when I stated that the units are single patient use, they can be used as long as the evaluation unit will calibrate but can not be autoclaved or sterilised - please see attached PDF for further details on disinfection.

I have requested a PPQ from our MD and will send this over as soon as it is available.

Should you have any further queries, please do not hesitate in contacting us.

Kind regards

Catrin Hollings
Sales & Marketing Co-ordinator
Viamed Ltd.

<http://www.viamed.co.uk>
Email: catrin.hollings@viamed.co.uk
Tel: +44 (0)1535 634 542
Fax: +44 (0)1535 635 582

Skype: viamed.catrin.hird

Twitter: twitter.com/ViamedLtd
Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

The information in this email is confidential and may be legally privileged.
It is intended solely for the addressee. Access to this email by anyone else is unauthorized.
If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542
