

Manometer Calibration record

Manometer Calibration Index Number	
Manometer Bar code	
Equipment Used	CE 077
Serial number Test Equipment	
Calibration Certificate Number	
Test Equipment Calibration Index Number	
Manometer @	Manometer Reading
5 mBar	4.8
10 mBar	9.9
20 mBar	20.0
30 mBar	29.8
40 mBar	39.8
50 mBar	49.9
60 mBar	59.7
70 mBar	69.9
80 mBar	79.9
Pass / Fail +/- 1 Mbar	
Tested By	PC
Tested On	20.3.14
Signed	for cur