

Manometer Calibration record

Manometer Calibration Index Number	
Manometer Bar code	
Equipment Used	CE 178
Serial number Test Equipment	
Calibration Certificate Number	
Test Equipment Calibration Index Number	
Manometer @	Manometer Reading
5 mBar	5.2
10 mBar	10.2
20 mBar	20.2
30 mBar	30.2
40 mBar	40.3
50 mBar	50.1
60 mBar	60.1
70 mBar	70.2
80 mBar	80.1
Pass / Fail +/- 1 Mbar	
Tested By	PC
Tested On	20.3.14
Signed	Dec 