

CERTIFICATE OF CALIBRATION

Issued by: RS Components Ltd

Date Issued: 27 Nov 2013

Certificate No.

1331638



RS Calibration

Calibration and Repair Service

DPN 175, Lammas Rd,
Weldon Industrial Est
Corby, Northants, NN17 9RS

Tel: 01536 405545
Fax: 01536 401590

Page 1 of 3 Pages

A handwritten signature in black ink that reads 'I. Philips'. Below the signature, the name 'Ian Philips' is printed in a smaller, black, sans-serif font.

Client

VIAMED LTD
KEIGHLEY
WEST YORKSHIRE
BD20 7DT

Instrument

Seaward Primetest 100 Portable Appliance Tester

Serial No.

Z28-0458

Client Reference

N/A

Procedure ID.

613.0573 Rev. P3

Date of Calibration

27 Nov 2013

Equipment Used to Carry Out Calibration

Transmille 3200 Electrical Test Equipment Calibrator
Transmille IEC Lead

Equipment ID.

Cal 960
Cal 3005

The measurements reported in this certificate were carried out using equipment whose values are traceable to national standards.

All procedures employed and results reported are in compliance with the requirements of the International Standard ISO/IEC 17025:2005.

The management controls of the RS Calibration Laboratory are registered under the British Standard BS EN ISO 9001 : 2008 No. RS 00362.

Uncertainties

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

This certificate reports recorded values for the instrument 'As Received'.



RS Components takes its environmental responsibilities very seriously and as such has printed this double sided document in black and white, on paper from sustainable sources.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate No.
1331638



Calibration and Repair Service

Page 3 of 3 Pages

Compliance to Specification

The specification published by the manufacturer and found in the instrument's handbook has been used to determine performance at the measured points.

Reported values not annotated.

The instrument complies with the stated specification, due allowance having been made for the uncertainty of measurement which carries no implication regarding the long term stability of the instrument.