

CERTIFICATE OF CALIBRATION

Issued by: RS Components Ltd

Date Issued: 14 Mar 2014

Certificate No.

1345023



RS Calibration

Calibration and Repair Service

Page 1 of 2 Pages

DPN 175, Lammas Rd,
Weldon Industrial Est
Corby, Northants, NN17 9RS

Tel: 01536 405545
Fax: 01536 401590

A handwritten signature in black ink that reads 'M.N. Tiney'. Below the signature, the name 'Martyn Tiney' is printed in a smaller, standard font.

Client	VIAMED LTD KEIGHLEY WEST YORKSHIRE BD20 7DT
Instrument	Digitron P200UL.is 0 to 100 mbar Digital Manometer
Serial No.	420720660
Client Reference	CE149
Procedure ID.	388.3555 Rev. P3
Date of Calibration	14 Mar 2014
Performance Status	Pass

Equipment Used to Carry Out Calibration	Equipment ID.
Ruska Deadweight Tester	Cal 711
Weights	Cal 713, 714

The measurements reported in this certificate were carried out using equipment whose values are traceable to national standards.

All procedures employed and results reported are in compliance with the requirements of the International Standard ISO/IEC 17025:2005.

The management controls of the RS Calibration Laboratory are registered under the British Standard BS EN ISO 9001 : 2008 No. RS 00362.

Uncertainties

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

This certificate reports recorded values for the instrument 'As Received'.



RS Components takes its environmental responsibilities very seriously and as such has printed this double sided document in black and white, on paper from sustainable sources.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate No.
1345023



Calibration and Repair Service

Page 2 of 2 Pages

The ambient temperature and relative humidity throughout the calibration were $(20 \pm 2) ^\circ\text{C}$ and $(50 \pm 20) \% \text{RH}$ respectively.

The instrument was calibrated using a dead weight pressure standard, applying a set mass to achieve nominal values of gauge pressure and recording the instrument's displayed values.

Prior to calibration the instrument was exercised to full scale pressure and then returned to zero.

The pressure medium used was nitrogen.

Range	Applied Value	Instrument Reading	Deviation as % of F.S.	L.S.D Stability
100 mbar	0 mbar	0.00 mbar	0.00	0
	30 mbar	29.9 mbar	0.10	0
	50 mbar	50.0 mbar	0.00	0
	75 mbar	75.1 mbar	0.10	0
	100 mbar	100.2 mbar	0.20	0

Measurement uncertainty of the above recorded values: $\pm (0.01\% + 1 \text{ L.S.D})$

END OF CALIBRATION

CALIBRATED BY:- MNT

Compliance to Specification

The specification published by the manufacturer and found in the instrument's handbook has been used to determine performance at the measured points.

Reported values

The uncertainties quoted refer to the applied values, which include any identified contribution of the instrument under test and not to the ability of the instrument to maintain its calibration.