

## DVM Calibration record

|                                         |                                                                                       |
|-----------------------------------------|---------------------------------------------------------------------------------------|
| DVM Calibration Index Number            | CE058                                                                                 |
| DVM Bar code                            |                                                                                       |
| Equipment Used                          | Time 1030 MicroCal                                                                    |
| Serial number Test Equipment            | 6839k6                                                                                |
| Calibration Certificate Number          | <del>1098426</del> 1345344                                                            |
| Test Equipment Calibration Index Number | CE076 (Barcode 102060)                                                                |
|                                         |                                                                                       |
| Micro Cal @                             | DVM Reading                                                                           |
| 0mV +/- 0.1                             | 0.0                                                                                   |
| 5mV +/- 0.1                             | 4.9                                                                                   |
| 10mV +/- 0.1                            | 9.9                                                                                   |
| 15mV +/- 0.1                            | 14.9                                                                                  |
| 20mV +/- 0.1                            | 19.9                                                                                  |
| 25mV +/- 0.1                            | 25.0                                                                                  |
| 30mV +/- 0.1                            | 30.0                                                                                  |
| 35mV +/- 0.1                            | 35.0                                                                                  |
| 40mv +/- 0.1                            | 40.0                                                                                  |
| Pass / Fail                             | Pass                                                                                  |
| Tested By                               | DLAM83                                                                                |
| Tested On                               | 15/10/14                                                                              |
| Signed                                  |  |