

DVM Calibration record

DVM Calibration Index Number	CE11)
DVM Bar code	
Equipment Used	Time 1030 MicroCal
Serial number Test Equipment	6839k6
Calibration Certificate Number	1098426 1345344
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
Micro Cal @	DVM Reading
0mV +/- 0.1	-0.1
5mV +/- 0.1	5.0
10mV +/- 0.1	10.0
15mV +/- 0.1	15.0
20mV +/- 0.1	20.0
25mV +/- 0.1	25.0
30mV +/- 0.1	30.0
35mV +/- 0.1	35.0
40mv +/- 0.1	40.0
Pass / Fail	Pass
Tested By	DLGMB
Tested On	15/10/14
Signed	