

Tektronix (China) Co. Ltd.  
1227 Chuan Qiao Road  
Pudong New District, Shanghai, China 201206

**Tektronix®**

TRACEABLE CALIBRATION  
TBS1052B, C013518  
Cal Date: 07-OCT-2014  
Cal Due: Lab: Tek  
**Tektronix®**

## CERTIFICATE OF TRACEABLE CALIBRATION

Certificate No: PC9KW9RJ7S

Manufacturer: Tektronix

Description: Oscilloscope; 50MHz, 2 channel

Cal Date: 07-OCT-2014

Revision: 00

Model: TBS1052B

Temperature: 24.4 °C

Serial No: C013518

Humidity: 47 %

Due Date: \_\_\_\_\_ \*

Date Placed In Service: \_\_\_\_\_ \*

\*Optional customer entry fields: The "Due Date" may be established by adding the calibration interval to the "Date Placed In Service." Recommended calibration intervals, for Tektronix products, are available on the Tektronix website at: <http://www.tek.com/service/product-support/>

### INSTRUMENT CONDITION:

Received: Not applicable, initial calibration

Returned: In tolerance, within published measurement specification

Tektronix certifies the above instrument has been calibrated using standards traceable to the PRC National Institute of Metrology (NIM) and/or other National Metrology Institutes (NIST, NPL, PTB) that are linked to the international system of units (SI). The policies and procedures used for the calibration of this product comply with ISO/IEC 17025:2005.

This certificate shall not be reproduced, except in full, without the written approval of the calibration facility.

### CALIBRATION PROCEDURE:

MANIFEST:Product\_Oscilloscopes\_TDSTBSPS1K2K-9100-2012\_Full VERSION:40

### CALIBRATION EQUIPMENT USED:

MANUFACTURER/MODEL	MODEL DESCRIPTION	ID NUMBER	DUE DATE
Keithley 2000	Multimeter, 6 1/2 digit	E126336	04-Apr-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF19-01	20-Nov-2014
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF19-02	20-Nov-2014
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF19-03	20-Nov-2014
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF19-04	20-Nov-2014
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF19-05	20-Nov-2014
Wavetek 9100	Calibrator	CT0016	08-May-2015

Issued By:

Quality Director: Cui, Hu Wa

Certified By: Zihai Zhang

Date Issued: 07-OCT-2014

ISO 9001 Registered Quality System

TBS1052B C013518

Page 1 of 1

001-1387-01

