

Ultra Max O₂ Oxygen analyzer

Contents list.

Monitor

Silicone cover

Yes/No

User manual

2x AA batteries.

Oxygen test.

Calibration Verification

98.5% to 101.5% reading = Free type
single line

Test flow at 8 lpm ± 0.2 Pass/Fail

" " " 5 lpm ± 0.2 P/F

" " " 2 lpm ± 0.2 P/F

Pressure test 0.8 to 1.0 psi P/F

Air test

19.4% to 22.4% reading Free type
single line