

DVM Calibration record

DVM Calibration Index Number	CE055
DVM Bar code	102039
Equipment Used	Time 1030 MicroCal
Serial number Test Equipment	6839k6
Calibration Certificate Number	1098426
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
Micro Cal @	DVM Reading
0mV +/- 0.1	0.0
5mV +/- 0.1	5.0
10mV +/- 0.1	10.0
15mV +/- 0.1	15.0
20mV +/- 0.1	20.0
25mV +/- 0.1	25.0
30mV +/- 0.1	30.0
35mV +/- 0.1	35.0
40mv +/- 0.1	40.0
Pass / Fail	PASS
Tested By	M. Green
Tested On	12-6-15
Signed	M. Green