

QA

0310030 Tom Thumb Infant Resuscitator.

Barcode ID: 597335

Serial Number 040900

Carried out by: Philip Crossley

03/Aug/2017 11:51 am

Performance evaluation: Pass

Signed: _____

Name: _____

Position: _____

Date: _____

QA 0310030 Serial Number 040900 Barcode ID: 597335
Carried out by: Philip Crossley 03/Aug/2017 11:51 am

[1]2173 Customer Requested Maximum Limit

[2]804 Serial Number
040900

[3]805 Calibration Date
03 / 08 / 2017

[4]806 Snoop Test No Leaks Pass
Yes

[5]807 Adjustable Value @ 10lpm Minimum is Max of 8cmH2O
Yes

[6]2835 Adjustable Value @ 10lpm 42.1 +/- 1 mbar
42.6

[7]809 Precision Value @ 10lpm Maximum 44.1 mbar +/- 1
44.1

[8]810 Pressure Gauge @ 10 lpm Pressure 30cmH2O +/- 1 (28.4 - 30.4 mbar)
30.1

[9]811 Gauge cannot be removed without Tools PASS
Yes

[10]812 Gauge is Straight PASS
Yes

[11]813 Labels are attached CE Label
Yes

[12]814 Labels are attached Serial No. Label
Yes

[13]815 Labels are attached Tom Thumb Label
Yes

[14]816 Built By
Philip Crossley

[15]871 Calibrated Manometer ID
102062

[16]2009 Customers Asset Number (service only)