

mV Reading in Air	PPO2 Reading @ 100% O2	mV reading in 100% O2	PPO2 Reading @ 100% O2 - after calibration	mV Reading @ 100% O2 - after calibration	PPO2 reading in air	mV reading in air	PO2 Reading @ 100% O2	0.5 BarG mV Reading @ 100% O2	0.5 BarG PO2 Reading @ 100% O2 (1.40 - 1.60)	1 BarG mV Reading @ 100% O2	1 BarG PO2 Reading @ 100% O2 (1.90 - 2.10)	1.3 BarG mV Reading @ 100% O2	1.3 BarG PO2 Reading @ 100% O2 (2.20 - 2.40)	1.5 BarG mV Reading @ 100% O2	1.5 BarG PO2 Reading @ 100% O2 (2.40 2.60)	mV Reading @100% O2	PO2 Reading @100% O2	mV reading in air	PPO2 reading in air
11	2.344	59	1.000	25.0	0.208	5.2	0.996	37.2	1.488	49.3	1.972	56.6	2.261	61.4	2.454	24.9	0.996	5.2	0.208
10.3	2.285	57.3	1.000	25.0	0.208	5.2	0.996	37.2	1.488	49.3	1.970	56.4	2.259	61.3	2.451	24.9	0.996	5.2	0.208
10.5	2.273	56.9	1.000	25.0	0.209	5.2	0.996	37.2	1.488	49.2	1.970	56.5	2.260	61.3	2.454	24.9	0.995	5.2	0.207
10.8	1.973	49.6	0.999	25.0	0.208	5.2	1.000	37.4	1.493	49.4	1.976	56.7	2.268	61.4	2.455	25.1	1.002	5.2	0.208
11	1.997	50	0.999	25.0	0.208	5.2	1.002	37.4	1.495	49.4	1.978	56.7	2.269	61.4	2.459	25.1	1.002	5.2	0.208
11.4	2.076	51.9	0.999	25.0	0.208	5.2	1.000	37.3	1.493	49.4	1.975	56.6	2.268	61.4	2.456	25.1	1.002	5.2	0.208
10.9	2.082	52.3	1.001	25.0	0.206	5.1	1.000	37.3	1.490	49.5	1.979	56.8	2.270	61.6	2.465	25.0	1.001	5.2	0.209
10.6	2.030	50.8	1.001	25.0	0.206	5.1	1.000	37.2	1.489	49.4	1.977	56.7	2.267	61.6	2.460	25.0	1.001	5.2	0.208
10.4	1.984	49.6	1.001	25.0	0.206	5.1	1.000	37.2	1.490	49.4	1.977	56.7	2.269	61.7	2.463	25.0	1.002	5.2	0.209
10.9	2.040	51.2	1.000	25.0	0.206	5.2	0.998	37.2	1.486	49.3	1.971	56.5	2.258	61.4	2.454	24.9	0.996	5.2	0.207
10.8	2.004	50.2	1.000	25.0	0.207	5.2	0.998	37.2	1.489	49.3	1.975	56.6	2.263	61.5	2.460	24.9	0.996	5.2	0.207
11.1	2.058	51.4	1.000	25.0	0.207	5.2	0.998	37.2	1.489	49.3	1.974	56.5	2.263	61.4	2.459	24.9	0.996	5.2	0.208
11	2.018	50.7	1.000	25.0	0.208	5.2	0.996	37.2	1.487	49.3	1.969	56.5	2.259	61.2	2.447	24.9	0.995	5.2	0.208
10.8	1.960	49.1	1.000	25.0	0.208	5.2	0.996	37.2	1.486	49.2	1.969	56.4	2.259	61.1	2.447	24.9	0.996	5.2	0.208
10.8	1.999	50	1.000	25.0	0.208	5.2	0.997	37.1	1.487	49.2	1.967	56.4	2.256	61.1	2.445	24.9	0.995	5.2	0.208
10.4	1.965	49.3	1.000	25.1	0.207	5.2	0.999	37.2	1.486	49.2	1.966	56.5	2.258	61.3	2.449	24.9	0.995	5.2	0.208
10.9	2.073	51.9	1.000	25.0	0.207	5.2	0.999	37.1	1.484	49.1	1.962	56.3	2.251	61.1	2.445	24.9	0.998	5.2	0.209
10.4	1.992	49.8	1.000	25.1	0.207	5.2	0.996	37.2	1.187	49.2	1.967	56.5	2.259	61.3	2.453	24.9	0.997	5.2	0.208
10.8	2.014	50.4	1.001	25.1	0.208	5.2	0.997	37.4	1.483	49.6	1.982	56.6	2.260	61.5	2.463	24.9	0.997	5.2	0.206
10.9	2.034	50.8	1.001	25.1	0.208	5.2	0.997	37.1	1.484	49.6	1.983	56.5	2.260	61.5	2.464	24.9	0.997	5.2	0.206
10.9	1.932	48.2	1.001	25.1	0.208	5.2	0.997	37.1	1.481	49.5	1.978	56.5	2.255	61.4	2.459	25.0	0.995	5.2	0.207
10.9	2.096	52.7	1.000	25.0	0.209	5.2	0.997	37.3	1.491	49.3	1.973	56.9	2.275	61.6	2.467	25.0	0.998	5.2	0.207
10.3	1.980	49.6	1.000	25.0	0.207	5.2	0.997	37.3	1.492	49.3	1.975	56.9	2.277	61.7	2.468	24.9	0.999	5.2	0.208
10.6	2.049	51.3	1.000	25.0	0.207	5.2	0.997	37.2	1.488	49.2	1.972	56.9	2.270	61.5	2.463	24.9	0.996	5.2	0.207
9.8	1.889	47.3	1.000	25.1	0.210	5.3	0.997	37.5	1.498	50.1	1.998	57.4	2.291	62.3	2.488	25.1	1.001	5.3	0.208
11.8	2.276	56.9	1.000	25.0	0.210	5.3	0.996	37.6	1.502	50.1	2.003	57.5	2.298	62.4	2.496	25.2	1.005	5.3	0.209
9.7	1.861	46.6	1.000	25.0	0.210	5.3	0.999	37.5	1.497	49.9	1.996	57.3	2.290	62.2	2.486	25.1	1.001	5.2	0.208
10.3	2.003	50.4	1.000	25.0	0.210	5.2	0.997	37.4	1.498	49.7	1.988	56.8	2.274	62	2.478	24.9	0.996	5.2	0.208
10.5	2.005	50.4	1.000	25.0	0.210	5.2	1.001	37.3	1.488	49.5	1.980	56.4	2.265	61.8	2.470	25.0	1.000	5.2	0.208
10.3	2.016	50.5	1.000	25.0	0.209	5.2	0.997	37.4	1.495	49.7	1.989	56.9	2.276	62	2.480	24.8	0.997	5.2	0.208
10.3	1.970	49.5	1.000	25.0	0.207	5.2	0.996	37.1	1.480	49.4	1.972	56.8	2.263	61.7	2.464	25.0	0.999	5.2	0.208
10.5	2.000	50.2	1.000	25.0	0.207	5.2	0.997	37.2	1.484	49.5	1.978	56.7	2.270	61.8	2.471	25.0	1.001	5.2	0.208
10.3	1.963	49.2	1.000	25.0	0.209	5.2	0.997	37.2	1.489	49.5	1.978	56.7	2.270	61.8	2.471	25.1	1.002	5.2	0.208
8.3	2.030	51.3	1.000	25.1	0.206	5.2	0.997	37.3	1.486	49.6	1.979	56.8	2.267	61.8	2.464	24.9	0.995	5.2	0.206
8.3	2.038	51.2	1.000	25.1	0.206	5.2	1.000	37.3	1.490	49.5	1.979	56.7	2.264	61.6	2.461	24.9	0.997	5.2	0.206
8.7	2.090	51.4	1.000	25.1	0.206	5.2	0.998	37.3	1.490	49.6	1.983	56.9	2.272	61.8	2.470	25.0	0.997	5.2	0.206
9.9	1.902	48	1.001	25.0	0.208	5.2	0.998	37.2	1.486	49.5	1.976	56.8	2.268	61.5	2.458	24.9	0.997	5.2	0.208
9.8	1.890	47.4	1.001	25.0	0.208	5.2	0.998	37.1	1.484	49.2	1.969	56.6	2.262	61.3	2.450	25.0	1.000	5.2	0.208
10.2	1.976	49.5	1.001	25.0	0.208	5.2	0.998	37.2	1.488	49.5	1.980	56.8	2.273	61.5	2.460	25.0	1.000	5.2	0.208