

TEST CERTIFICATE

RESULT :- PASS

Part Number :- C5288

Serial Number :- RMA

Customer Details :- VIAMED - V1000

Time : 09:00:41

Date : 21/08/2017

Passes : 1000000

Fails : 0

FailsUp : 0

FailsDown : 0

Starttime : 15/08/2017 15:47:21

Testtime : 86:44:47

Resistance RMin 7

RMax 12

Test Specification

TimeUp mS: 100

TimeDown mS: 200

Test Cycles: 1000000

Resistance Spec: 100

Force: 0 gr

Range: 1K 5mA

Test Engineer DARREN RHODES

Signature 

Date 21/08/2017