

DVM Calibration record

CE 169

Equipment description if new	FLUKE 187
Serial Number if new	90050368 (DEC 2005)
DVM Calibration Index Number	CE 165
DVM Bar code	
Equipment Used	Time 1030 MicroCal
Serial number Test Equipment	6839k6
Calibration Certificate Number	1098426
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
Micro Cal @	DVM Reading
0mV	0
5mV	5.002
10mV	9.997
15mV	14.999
20mV	20.007
25mV	25.039
30mV	30.050
35mV	35.047
40mV 50mV	50.013
Pass / Fail	
Tested By JS Lamb	
Tested On 10/06/08	
Signed JS Lamb	

EXTRA READINGS

100 mV	99.94
500 mV	499.8
1.0 V	999.4

Next Calibration Required 2010.

