



Queensland Government

Department of Industrial Relations

Electrical Safety Act 2002

Certificate of Approval for an Electrical Article

Registration No: Q061389

This is to certify that the Regulator has approved the electrical article described hereunder.

Registered Declarant: FRIWO Geratebau GmbH
Von-Liebig-Strasse 11
D-48346 Ostbevern - GERMANY

DETAILS OF ARTICLE

Article:	Power Supply Switch Mode Power Supply for IT or Medical use
Trade Name:	FWGB
Catalogue/Model/Type Number:	FW7660O/xx SERIES
Marking Details:	Input: 100-240Vac 50-60Hz 250mA Output: 5-24Vdc 1600-420mA
Reference Number:	061389
Relevant Standard:	AS/NZS60950.1: 2003
Date of Registration:	02 March 2007
Expiry Date:	01 March 2012

**Director – Equipment Safety
Electrical Safety Office**

13/03/2007



Queensland Government

Department of Industrial Relations

Electrical Safety Act 2002

Attachment to Certificate of Approval for an Electrical Article

Part A

Approval Number:	Q061389
Modification to:	Power Supply Switch Mode Power Supply for IT or Medical use
Trade Name:	FWGB
Catalogue/Model/Type Number:	FW7660O/XX SERIES
Marking Details:	Input: 100-240Vac 50-60Hz 250mA Output: 5-24Vdc 1600-420mA

Part B

Date of Modification: 05 March 2007

Details of Modification:

- Where "M" in model name stands for Medical use and "O" in Model name stands for Office use.
- Where xx can be 05-24 to indicate output voltage 5.0-24V DC output.
- Alternative tradename "FWHK"

Additional Models

FW7660M/XX

Description

Similar to FW7660O/xx except for:

- model number,
- medical applications in lieu of office use; and
- additional compliance to AS/NZS3200.10:1988.


Director – Equipment Safety
Electrical Safety Office
15/3/2007



Queensland Government
Department of Industrial Relations