
Ceratherm 600-3 Images: FAO Marketing

Jonas Boss <Jonas.Boss@ewimed.com>

23 July 2026 at 15:43

To: "catrin.hollings@viamed.co.uk" <catrin.hollings@viamed.co.uk>

Cc: Daniel Zaragoza <Daniel.Zaragoza@ewimed.com>

Hi Catrin,

Thank you for updating the images and replacing the logo. The visual amendments and the ewimed logo look good.

Before the brochure is used, however, could you please make the following technical and regulatory corrections based on the current Ceratherm 600-3 product specifications:

Page 1 – Intended purpose

The current wording states:

“For maintaining the body temperature of infants, children and adults”

Please replace this with:

“For maintaining the body temperature of newborns, infants and young children and for pre-warming care surfaces.”

The reference to adults should be removed, and the pre-warming of care surfaces should be added in accordance with the intended purpose.

Page 2 – Model designation

For the version with double hinged arm and wall mounting bracket, the model is currently stated as:

“521A-60022-3-OND”

Our documentation lists the model as:

“521A-60022-3-CH(-EX)”

Could you please check whether “-OND” is a Viamed-specific designation and intentional? Otherwise, it should be corrected accordingly.

Page 4 – Technical data

Please amend the following specifications:

- **Lighting**

Current: “LED lamp 12 V / 1 W”

Correct: **“LED / 1 × 3 W / colour temperature 4000 K”**

- **Radiator performance**

Current: “4 steps, 0–99%”

Correct: **“4 programmable steps, 0–98%”**

- **Notified Body / certification**

Current: “IEC 601-2 TÜV/CE”

Please replace this with:

“Notified Body: TÜV SÜD Product Service GmbH, CE 0123”

- **Pollution degree**

Please use the term:

“Pollution degree: 2 according to EN 60601-1”

- **Applied standards**

Please replace the standards currently listed with:

EN 60601-1:2006 + A1:2013

EN 60601-2-21:2009 + A11:2011 + A1:2016

EN 60601-1-2:2015

- **Warmer dimensions**

Current: “55 cm (L) × 21 cm (W) × 9 cm (H)”

Correct: **“54 cm (L) × 27 cm (W) × 10.5 cm (H)”**

There is also a small editorial error in the description on page 4:

“from to 10–100% illumination”

This should read:

“from 10–100% illumination”

Once these points have been corrected, please send us the revised version for final review before publication.

We would also appreciate receiving the image files in the highest available resolution, as they may be useful for our own marketing materials.

I would also like to inform you about some organisational changes at ewimed Switzerland AG. Daniel Urfer will retire at the end of July 2026.

I have taken over his responsibilities for quality, regulatory and technical matters and now act as COO, Quality Manager and PRRC.

Daniel Zaragoza, copied on this email, is our new CEO Switzerland and Executive Director and will be responsible for commercial matters relating to the Ceratherm product range.

For future Ceratherm correspondence, please therefore include Daniel Zaragoza and me instead of Daniel Urfer.

Thank you very much for your support.

Kind regards,

Jonas Boss

COO | Quality Manager & PRRC


ewimed Switzerland AG
Freiburgstrasse 453
CH-3018 Bern

Mobile +41 79 293 92 84
Phone +41 31 958 66 66
E-Mail Jonas.Boss@ewimed.com

Allgemeine Geschäftsbedingungen
Conditions générales de vente
Termini e condizioni generali
General terms and conditions

[Quoted text hidden]



Radiant Warmers 2026-07-23.pdf
10029K