

Calibration Verification

The UltraMaxO2 Oxygen Analyser has been tested for accuracy using medical grade Oxygen and air supplied by BOC.

Medical test gases used:

Medical Oxygen (BOC)	PL 0735/5000
Purity	99.5% (min)
Batch No.	15962234

Medical Air (BOC)	PL 0735/5002R
Purity	21.0%
Batch No.	15927686

Serial number of the UltraMaxO2	GK74511005
REF	R221P11-001

Oxygen input:	UltraMaxO2 reading:
99.5%	100.3%

Air input	UltraMaxO2 reading:
21.0%	20.8%

Air Flow Test:	Pass
Pressure Test:	Pass

Certificate Number	1777366524
--------------------	------------

Date: 28th April 2026

Tested by: P Crossley

Signed: