

Service Repair Sheet SRS69468

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Stinder Bassi
Queen Elizabeth University Hospital
Medical Physics Department
Specialist Clinical Technologist
01414523030
01412011100
00001853
stinder.bassi2@nhs.scot

13/Jul/2026
Robert Connor
Viamed
Quote

VIAMEDclean	
Goods In Only	☒
Decontamination certificate provided by customer	
Cleaned by Viamed, if no declaration certificate from customer	☒
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 08/Jul/2026 Emily Hanson

08/Jul/2026 Emily Hanson

VM-2500-S C02/SpO2 monitor S/N 60940148 that has an issue with the real time clock. The device will not store the correct time when entered.

13/Jul/2026 Robert Connor

Received 1 x VM-2500-S s/n 60940148, with silicone cover.

Ready For quote

CGNEON 24.7.26

Repair Complete Signed

CGNEON 31.7.26

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN38916	Capnograph	4410520	60940148	N

44400000 x 1 @ 0.5hr = £49.35
S/n, SES, SRN

UPS x 1 @ £12.