

DVM Calibration record

1	DVM Calibration Index Number	CE206
2	DVM Bar code	1288063
3	Equipment Used	MICROCAL 1030
4	Serial number Test Equipment	3412J25
5	Calibration Certificate Number	1955273
6	Test Equipment Calibration Index Number	CE223
	Micro Cal @	DVM Reading
7	0mV +/- 0.1	0.0
8	5mV +/- 0.1	5.0
9	10mV +/- 0.1	10.0
10	15mV +/- 0.1	15.0
11	20mV +/- 0.1	20.0
12	25mV +/- 0.1	25.0
13	30mV +/- 0.1	30.0
14	35mV +/- 0.1	35.0
15	40mv +/- 0.1	40.0
16	Pass / Fail	PASS
17	Tested By	PHIL CROSSLEY
18	Tested On	6.8.26
19	Signed	

1. Enter the details of the unit being tested in boxes 1 – 2
2. Enter the details of the calibrated Micro Cal unit being used to test against the equipment. In to boxes 3 -6
Ensure the calibrated unit is still inside its calibration period before you begin.
3. set the micro cal to the required settings (7 – 15) and enter the readings on the DVM in the Boxes
4. Assuming the units passes the tests, date and sign the form (17 – 19). Scan the form and upload to the issue requesting calibration and complete the Issue.
If the unit fails the test return the Unit to RS for calibration