



## **EC DECLARATION OF CONFORMITY**

Manufacturer: BeiJing Choice Electronic Technology CO., LTD.

Address: Bailangyuan Building B 1127-1128, Fuxing Road A36, Beijing,  
P.R.CHINA

European representative: Shanghai International Holding Corp. GmbH (Europe)

Address: Eiffestrasse 80, 20537 Hamburg Germany

Product: Pulse Oximeter

Model Code: MD300C2

Classification: Class II a by Rule 10 of Annex IX, MDD93/42/ECC

Conformity of assessment route: Annex II .3, MDD93/42/ECC

We herewith declare that the above mentioned products meet the provisions of the following EC Council Directives and Standards. All Supporting documentation are retained at the premises of the manufacturer.

### **Directives**

General applicable directives:

Medical Device Directive: Council Directive 93/42/ECC of 14 June 1993 concerning medical devices (MDD93/42/ECC).

Standards:

Harmonized Standards (published in the Official journal of the European Communities) applicable to this product are:

ISO 13485:2003, ISO14971:2007, ISO10993-1:2003, ISO10993-5:1999,  
ISO10993-10:2002/A1:2006, IEC60601-1:1988\_A1:1991\_A2:1995,  
EN60601-1-2:2001, ISO9919:2005, EN1041:1998, EN980:2003.

Notified Body: TUV SUD PRODUCT SERVICE GMBH

Report No. : BJ890104

Valid until: 2014-05-21

Date CE mark was first affixed: 01 06 2006

Signature

